



AB GERMA®

EU Declaration of conformity no. 230710-001

Product Name:	Motion Restriction Pillow MRP Black FW
Intended use:	The pillow is intended to be used in prehospital and hospital environment, by trained emergency and hospital personnel to stabilise patient's head during transport. It can be used either as a comfort pillow or a vacuum pillow. The pillow fits most ambulance stretchers.
SRN: Basic UDI-DI: UDI DI: Germa Article No:	SE-MF-000003932 735001959P04POSPILLQC 07350019596210 15000000115
Manufacturer: Visiting address: Phone: Email: Web:	AB Germa Industrigatan 54-56, SE-29136 Kristianstad +46 (0)44 123030 info@germa.se www.germa.se
Product class:	Class I according to rule 1 in Annex VIII in MDR 2017/745
Conformity procedure:	Self-certification according to Annex IV in MDR 2017/745
Identification:	All products with serial numbers issued from; LOT number: 517200 Date: 2021-03-23

Declaration statement;

This EU declaration of conformity is issued under the sole responsibility of the manufacturer AB Germa. The devices covered by this declaration is in conformity with the requirements in the European Medical Device Regulation 2017/745.

Signed in Kristianstad, Sweden on behalf of AB Germa by:

Position: Managing Director
Name: Patrik Thornström
Date: 2023-07-10

Sign: _____

Postadress/Postal address

AB GERMA
Box 2127
SE-291 02 KRISTIANSTAD
SWEDEN

Adress/Goods address

Industrigatan 54-56
SE-291 36 Kristianstad

Telefon/Phone

+46 (0)44-12 30 30

E-mail/Web page

info@germa.se
www.germa.se

VAT No.

SE556086941301

Bankgiro
Bank account

481-4125