

## EC Declaration of conformity no.: 190310-003

| Product Name:     | Rescue Light All Body Splint  |
|-------------------|---|
| Germa Article No: | 23271513000   |
| Description:      | Rescue Light All Body Splint with belt system intended for motion restriction of the patient. Designed to be used in combination with Basket Stretcher. Vacuum used to safely make a rigid immobilization on the patient. Restraint system consist of four (4) colour-coded restraints, upper part chest and mid body/ hip and leg. Fixation on the mattress at dedicated fixation points and Fix lock closing of belts that always shall be used to secure the patient. Professional medical trained rescuers shall use this mattress.  Read User Manual for repair instructions |
| Accessory:        |   |

Identification: All products with serial numbers issued after LOT number: 030704, date: 2003-07-04 (yyyy-mm-dd).

We the manufacturer hereby declare that this product is in conformity with the requirements in:

- Annex I and Annex VII and Annex IX in Council directive 93/42/EEC. Equipment class: I, non-sterile, no measuring function.
- Swedish Medical Devices Act 1993:584

Signed in Kristianstad, Sweden on behalf of AB Germa by:

Position:

Managing Director

Name:

Björn Holmqvist

Date:

10/03-2019

Sign:

Postadress Adress Telefon Telefax Organisationsnummer. Postgiro Box 2127 F-skattebevis S-291 02 KRISTIANSTAD Industrigatan 54-56 **SWEDEN** KRISTIANSTAD +46 (0)44-123030 +46 (0)44-103179 556086-9413 481-4125 21 68 22-7 e-mail: info@germa.se